

K121446



BacT/ALERT® PF Plus Culture Bottle

510(k) SUMMARY

JAN 25 2013

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR 807.92.

Name of device

BacT/ALERT® PF Plus

Device Identification

Trade Name: BacT/ALERT® PF Plus Culture Bottle

Classification Name: Blood Culturing System, Microbiology

Product Code: MDB

Regulation: 21CFR866.2560, microbial growth monitor

Device Class: Class 1, not exempt from premarket notification per 21CFR807.81

Premarket Notification Submitter

Company Name: bioMérieux, Inc.
Company Address: 100 Rodolphe Street
Durham, NC 27712

Contact: Elizabeth Landon, Staff Regulatory Affairs Specialist
Telephone #: 919-620-2329
FAX #: 919-620-2548

Alternate Contact: Jocelyn Jennings, Senior Manager, Regulatory Affairs
Telephone #: 919-620-2894
FAX #: 919-620-2548

Preparation Date: May 11, 2012

Intended Use of the Device

BacT/ALERT® PF Plus Culture Bottles are used with the BacT/ALERT® Microbial Detection System in qualitative procedures for recovery and detection of aerobic and facultative anaerobic microorganisms (bacteria and yeast) from blood.

Description of the Device

The new reagent (BacT/ALERT PF Plus Culture Bottle) is an improvement upon the cleared charcoal formulation reagent (BacT/ALERT PF Culture Bottle). The BacT/ALERT PF Culture Bottles are used with the BacT/ALERT Microbial Detection Systems in qualitative procedures for recovery and detection of microorganisms from blood.

The predicate BacT/ALERT PF Culture Bottle contains charcoal, for its antimicrobial neutralization properties, in a complex growth medium. Charcoal is eliminated in the proposed BacT/ALERT PF Plus Culture Bottle, and is replaced with two types of adsorbent resins in a complex growth medium. The proposed BacT/ALERT PF Plus Culture Bottle is optimized to increase antimicrobial neutralization properties, and to increase the clarity of Gram stains in comparison to the predicate BacT/ALERT PF Culture Bottle.

The BacT/ALERT Microbial Detection System provides both a microbial detection system and a culture medium bottle with suitable nutritional and environmental conditions for microorganisms commonly encountered in blood (adult and pediatric bottles) or other normally sterile body fluid samples (except urine) (adult bottles only) taken from a patient suspected of having bacteremia/fungemia. An inoculated bottle is placed into the instrument where it is incubated and continuously monitored for the presence of microorganisms that will grow in the BacT/ALERT bottles.

The BacT/ALERT Microbial Detection System utilizes a colorimetric sensor and reflected light to monitor the presence and production of carbon dioxide (CO₂) that is dissolved in the culture medium. If microorganisms are present in the test sample, carbon dioxide is produced as the microorganisms metabolize the substrates in the culture medium. When growth of the microorganisms produces CO₂, the color of the gas-permeable sensor installed in the bottom of each culture bottle changes from blue-green to yellow. The lighter color results in an increase of reflectance units monitored by the system. Bottle reflectance is monitored and recorded by the instrument every 10 minutes.

SUBSTANTIAL EQUIVALENCE INFORMATION

Predicate device name(s)

BacT/ALERT® PF Culture Bottle

Predicate device 510(k) number(s)

K020923

Comparison with predicate

The BacT/ALERT PF Plus Culture Bottle is claimed substantially equivalent to the BacT/ALERT PF Culture Bottle (K020923).

Table 1. Similarities and differences between the tests are outlined below:

Culture Bottle Characteristics: Changes versus K020923	
Specimen Sampling and Handling	unchanged
Assay Types	unchanged
Reaction Types	unchanged
Calibration	unchanged
Quality Control (by Operator)	unchanged
Principles of Operation	unchanged
Firmware	<p>No changes to firmware occurred.</p> <p>The structure of the firmware algorithm remain unchanged.</p> <p>The knowledge base specifications utilized by the firmware included changes to the initial value threshold variables used by the firmware algorithm.</p> <p>Variables related to controlling barcode recognition were adjusted to enable recognition of the new bottle type.</p>

Performance Characteristics

Analytical Testing

Analytical Sensitivity: Growth Performance

Data represent results from in-house seeded studies with blood obtained from healthy human volunteers. Multiple strains were tested for each species at target inoculum levels of 125 CFU per bottle. The species listed are representatives of clinically prevalent organisms in blood cultures.

Table 2. Growth Performance Results

Microorganism	Blood			
	% Recovery (n)	Range CFU/bottle	Time to Detection (hours)	
			Mean	Range
<i>Staphylococcus aureus</i>	100.0 (30/30)	54 - 150	13.3	12.2 - 15.2
<i>Escherichia coli</i>	100.0 (30/30)	71 - 254	11.2	10.3 - 11.7
<i>Pseudomonas aeruginosa</i>	100.0 (12/12)	74 - 148	15.7	13.7 - 17.8
<i>Klebsiella pneumoniae</i>	100.0 (12/12)	89 - 123	11.3	10.6 - 12.3
<i>Candida albicans</i>	100.0 (30/30)	88 - 298	29.0	19.2 - 52.8
<i>Streptococcus pneumoniae</i>	100.0 (30/30)	3 - 260	13.8	10.8 - 16.5
<i>Staphylococcus epidermidis</i>	100.0 (12/12)	44 - 135	17.6	14.3 - 18.8
<i>Enterococcus faecalis</i>	100.0 (12/12)	63 - 259	11.6	11.0 - 12.2
<i>Enterococcus faecium</i>	100.0 (12/12)	25 - 120	12.8	11.3 - 14.4
<i>Enterobacter cloacae</i>	100.0 (12/12)	111 - 200	11.6	10.8 - 12.5
<i>Candida glabrata</i>	100.0 (12/12)	118 - 281	43.5	27.3 - 64.8
<i>Haemophilus influenzae</i>	100.0 (12/12)	105 - 266	14.4	12.0 - 16.8
<i>Proteus mirabilis</i>	100.0 (12/12)	36 - 213	12.5	11.3 - 14.6

Less than 100% detection was observed for some microorganisms, to include *Capnocytophaga ochracea*, *Cardiobacterium hominis*, *Eikenella corrodens*, *Haemophilus parainfluenzae*, *Granulicatella adiacens*, and *Helicobacter cinaedi*

Antimicrobial Neutralization

Neutralization of antimicrobials by adsorbent polymeric beads varies depending upon dosage level and timing of specimen collection. Internal studies have demonstrated that antimicrobials are effectively neutralized by the BacT/ALERT PF Plus medium. In these tests, antimicrobials were added in clinically relevant concentrations directly to culture bottles during inoculation with susceptible strains. The effectiveness of the antimicrobials was confirmed by parallel testing using a non-neutralizing medium as a control. Antimicrobials from the following categories were neutralized by the medium: penicillins, glycylicyclines, polyenes, macrolides, triazoles, echinocandins, cefazolin, cefoxitin, ceftaroline, aminoglycosides, fluoroquinolones, lincosamides, glycopeptides, and oxazolidinones. Antimicrobial neutralization was not achieved for ceftazidime or cefepime.

Less than complete neutralization was observed for ceftotaxime and ceftriaxone. Cefotaxime was neutralized at ranges of 50% peak serum level (PSL) to 2% PSL depending on the microorganism. Ceftriaxone was neutralized at ranges of 50% PSL to 1% PSL depending on the organism.

Potentially Interfering Substances

In-house seeded studies were conducted with plasma, blood, and blood clots. Aliquots of each of these fluids also received white blood cells at concentrations relevant to bacteremia in blood. Testing was conducted with and without microorganisms. These substances neither interfered with recovery and detection of organisms, nor did they generate false positive results in the absence of organisms.

Limit of Detection (LoD)

Data shows results from in-house seeded studies. A minimum of 30 replicates were tested per species. The data was generated using bottles at end of shelf life. Bottles inoculated with *H. influenzae* received 4 ml pooled human blood supplementation. At least 95% detection was achieved at LoD.

Table 3. Summary of LoD Data

Microorganism	Strain ID	LoD (CFU/bottle)
<i>Candida albicans</i>	ATCC 14053	6
<i>Enterobacter aerogenes</i>	ATCC 13048	8
<i>Enterococcus faecalis</i>	NCTC 12697	5
<i>Escherichia coli</i>	NCTC 12923	4
<i>Haemophilus influenzae</i>	ATCC 10211	6
<i>Klebsiella pneumoniae</i>	STL 104016	4
<i>Listeria monocytogenes</i>	ATCC 15313	6
<i>Pseudomonas aeruginosa</i>	NCTC 12924	4
<i>Salmonella enterica</i>	ATCC 14028	5
<i>Staphylococcus aureus</i>	NCTC 10788	5
<i>Streptococcus pneumoniae</i>	ATCC 6305	6

NOTE: 96.7% of the bottles were subcultured within 30 minutes of being declared positive. STL 104016 was sourced from internal culture collection.

Within-Laboratory Precision (Repeatability)

Data represents results from in-house seeded studies conducted on 12 days on multiple instruments by multiple operators. Organisms were grown in the presence of clinically relevant concentrations of antimicrobials to which they are susceptible. In this seeded study BacT/ALERT PF Plus bottles were subcultured at least 24 hours after being flagged positive by the instrument. A minimum of 108 replicates were tested for each organism/antimicrobial combination.

Table 4. Summary of the Within-Laboratory Precision Data

Sample Input		CFU/bottle (range)	% Recovery				Time to Detection (hours)	
Organism	Antimicrobial		Lot 1	Lot 2	Lot 3	Overall	Mean	Range
<i>C. albicans</i>	Fluconazole	140 - 364	100.0	100.0	100.0	100.0	26.0	22.8 - 31.3
<i>E. coli</i>	Amikacin	26 - 156	100.0	100.0	100.0	100.0	12.0	11.2 - 13.0
<i>K. pneumoniae</i>	Levofloxacin	108 - 170	100.0	100.0	100.0	100.0	13.4	11.7 - 15.2
<i>P. aeruginosa</i>	Piperacillin	80 - 148	100.0	97.2	100.0	99.1	19.2	17.4 - 24.1
<i>S. pneumoniae</i>	Penicillin G	9 - 505	100.0	100.0	100.0	100.0	13.2	11.6 - 15.5
<i>S. aureus</i>	Vancomycin	94 - 158	100.0	100.0	100.0	100.0	16.9	14.6 - 20.3

Reproducibility

Data represents results from seeded studies conducted at three sites using a target of 162 replicates per site on 3 days with a minimum of two operators per site. Reproducibility was evaluated on each of 9 organisms. Two organisms (*C. albicans* and *S. pneumoniae*) were prepared via serial dilution and the other 7 organisms were prepared using BioBalls. *C. albicans* and *S. pneumoniae* were seeded into the PF Plus bottle, at a target inoculum of 100 CFU/bottle, with an acceptable range of 30-300 CFU/bottle and the other 7 organisms at a target range of 1-17 CFU/bottle. Actual inoculum ranged from 6 CFU/bottle to 700 CFU/bottle for the 30-300 CFU/bottle range, and from 1 CFU/bottle to 270 CFU/bottle for the 1-17 CFU/bottle range. Percent recovery reflects positive flag by the instrument, gram stain/subculture consistent with the seeded organisms.

Table 5. Summary of Reproducibility Data

Sample Input	% Recovery				Time to Detection		Inoculum Ranges (CFU/Bottle)
	Site 1	Site 2	Site 3	Overall	Mean	Range	
<i>S. aureus</i>	100.0% (18/18)	87.5% (21/24)	100.0% (30/30)	95.8% (69/72)	15.6	14.6-16.7	2-11
<i>C. albicans</i>	100.0% (18/18)	83.3% (30/36)	100.0% (33/33)	93.1% (81/87)	36.6	24.6-76.8	14-700
<i>E. coli</i>	100.0% (27/27)	77.8% (21/27)	100.0% (30/30)	92.9% (78/84)	12.8	11.8-14.1	1-38
<i>P. aeruginosa</i>	100.0% (24/24)	75.0% (18/24)	97.0% (32/33)	91.4% (74/81)	18.4	17.1-21.1	1-11
<i>E. faecalis</i>	100.0% (18/18)	79.2% (19/24)	96.7% (29/30)	91.7% (66/72)	13.9	12.6-15.3	1-15
<i>E. aerogenes</i>	74.4% (29/39)	72.2% (26/36)	85.4% (41/48)	78.1% (96/123)	14.9	11.7-20.8	1-270*
<i>L. monocytogenes</i>	100.0% (18/18)	100.0% (24/24)	100.0% (30/30)	100.0% (72/72)	24.1	20.4-36.4	1-14
<i>S. enterica</i>	100.0% (24/24)	75.0% (18/24)	100.0% (33/33)	92.6% (75/81)	13.5	2.3-14.8	1-13
<i>S. pneumoniae</i>	100.0% (30/30)	100.0% (36/36)	100.0% (21/21)	100.0% (87/87)	14.2	11.6-18.9	6-500
Overall	95.4% (206/216) 95% CI: 91.7%, 97.8%	83.5% (213/255) 95% CI: 78.4%, 87.9%	96.9% (279/288) 95% CI: 94.2%, 98.6%	92.0% (698/759) 95% CI: 89.8%, 93.8%			

*Plate count of 270 CFU/bottle was arrived at by serial dilution.

The above data includes repeat testing performed as a result of laboratory errors at a single site (i.e. contaminated bottles/reagents, colony counts out of range, and site failure to change bottle status after positive instrument signal and positive subculture). Data excluding the laboratory errors demonstrated 100%

recovery with the exception of *E. aerogenes*, which exhibited 85.0% recovery for all sites combined.

Delayed Entry

Results from seeded studies using 11 species* at target concentrations 100 CFU per bottle (acceptable range of 30 to 300 CFU per bottle) were generated at three sites. Actual inoculum levels ranged from 35 CFU/bottle to 290 CFU/bottle. All bottles contained human blood from healthy volunteers and were held at specified temperatures and times prior to loading into the BacT/ALERT instrument. Percent recovery reflects positive flag by the instrument, gram stain/subculture consistent with the seeded organisms.

Table 6. Summary of Delayed Entry Data

Sample Input	Incubation Temperature (°C)	Hold Time (hours)	% Recovery	Time to Detection from Sample Inoculation (Hold Time + Instrument TTD in hours)	
				Mean	Range
Inoculated Test Bottles	Control	No delay	100.0% (459/459)	14.3	8.5 - 84.0
	2-8	48	98.6% (292/296)	63.7	57.5 - 103.2
	20-25	24	98.0% (291/297)	31.8	26.2 - 74.4
	20-25	36	91.9% (272/296)	41.8	38.0 - 70.5
	35-37	8	98.9% (454/459)	16.1	10.2 - 53.8
	35-37	24	56.6%*** (259/458)	28.3	26.0 - 74.4
Negative Controls	All conditions		0.5% (1/221)**	-	-

**Staphylococcus aureus*, *Candida albicans*, *Candida krusei*, *Escherichia coli*, *Klebsiella pneumoniae*, *Acinetobacter baumannii*, *Pseudomonas aeruginosa*, *Streptococcus pneumoniae*, *Enterococcus faecium*, *Haemophilus influenzae*, *Neisseria meningitidis*

**False positive observed during seeded study (1/221)

***CAUTION: Culture bottles held at 35 to 37°C for 24 hours or longer before loading may not detect microorganisms and should be subcultured.

Clinical Study Results (Blood Cultures)

A multi-center clinical study was conducted at three different geographic sites in the U.S. comparing the performance of the PF Plus and PF blood culture bottles for pediatric culture pairs that received blood volumes between 0.1 ml and 4 ml (compliant pairs). A total of 2188 bottle pairs were obtained from 1086 pediatric patients suspected of blood stream bacterial/yeast infections. Subcultures of both bottles were performed when either bottle in the set was determined to be positive by the BacT/ALERT system. A pair of bottles was determined to have a positive status if the subculture of either the PF Plus or PF bottle was positive. A culture bottle was determined to be a "True Positive" if the culture was flagged positive by the BacT/ALERT System and resulted in growth of the isolate upon subculture of this bottle. True positive rates were calculated for the PF Plus and PF culture bottles, and the ratio of PF Plus true positives to PF true positives was calculated to compare performance. Clinical isolates recovered were classified as significant, contaminant, or unknown based on determination by the clinical trial sites.

A total of 172 isolates were recovered from all compliant pediatric blood culture pairs with a positive status. There were a total of 145 bottle pairs that recovered at least 1 isolate by subculture of PF Plus or PF bottles. A total of 126 bottle pairs recovered a single isolate, 12 bottle pairs recovered two isolates, 6 bottle pairs recovered 3 isolates, and 1 bottle pair recovered 4 isolates. The total population reported in Table 7 comprises the 172 isolates recovered from positive bottle pairs and 2043 negative bottle pairs for a total of 2215 results. The BacT/ALERT PF Plus bottle detected a total of 140 isolates compared to the BacT/ALERT PF bottle that detected 128 isolates. Of the significant isolates, the BacT/ALERT PF Plus bottle detected a total of 91 isolates compared to the BacT/ALERT PF bottle that detected 77 isolates. One (1) false positive was identified by subculture of a positive BacT/ALERT PF Plus bottle and comprised 0.05% (1/2215) of the study population.

Table 7. All Compliant Pairs with Single and Multiple Isolates Combined (Blood Cultures)

Clinical Isolate Determination	BacT/ALERT PF Plus True Positives	% of BacT/ALERT PF Plus True Positives in Population	BacT/ALERT PF True Positives	% of BacT/ALERT PF True Positives in Population	Ratio of True Positives*
Significant	91	4.1% (91/2215)	77	3.5% (77/2215)	1.182
Contaminant	24	1.1% (24/2215)	29	1.3% (29/2215)	0.828
Unknown	25	1.1% (25/2215)	22	1.0% (22/2215)	1.136
<i>Total</i>	<i>140</i>	<i>6.3% (140/2215)</i>	<i>128</i>	<i>5.8% (128/2215)</i>	<i>1.094</i>

*Ninety six (96) isolates were detected by both PF Plus and PF, 44 isolates were detected only by PF Plus and 32 isolates were detected only by PF. The ratio of true positive rates for overall isolates was 1.094 (140/128) with a 95% CI of (0.954, 1.234).

Proposed labeling

The proposed labeling is complete.

Conclusion

The information in this premarket notification is complete and supports a substantial equivalence decision.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-002

bioMérieux, Inc
C/O Ms. Elizabeth Landon
Staff Regulatory Affairs Specialist
100 Rodolphe Street
Durham, NC 27712

JAN 25 2013

Re: k121446

Trade/Device Name: BacT/ALERT PF Plus Culture Bottle
Regulation Number: 21 CFR 2560
Regulation Name: Microbial Growth Monitor
Regulatory Class: Class I
Product Code: MDB
Dated: January 14, 2013
Received: January 15, 2013

Dear Ms. Landon:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements; including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostics and Radiological Health at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

Sally A. Hojvat

Sally A. Hojvat, M.Sc., Ph.D.

Director

Division of Microbiology Devices

Office of *In Vitro* Diagnostics and Radiological Health

Center for Devices and Radiological Health

Enclosure

INTENDED USE STATEMENT

510(k) Number (if known): K121446

Device Name: BacT/ALERT® PF Plus Culture Bottles

Intended Use:

BacT/ALERT® PF Plus Culture Bottles are used with the BacT/ALERT® Microbial Detection System in qualitative procedures for recovery and detection of aerobic and facultative anaerobic microorganisms (bacteria and yeast) from blood.


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


Division Sign-Off

Office of In Vitro Diagnostics and Radiological Health

Page 1 of 1

510(k) K121446